

TOWA MEDICAL INSTRUMENTS CO.,LTD.

Head office·factory: 1698-2, Oaza-Nakano, Nakanoshi, Nagano 383-0013, Japan
Telephone: +81-269-22-5181 Telefax: +81-269-22-5184
Tokyo office: 2-10-10, Yushima, Bunkyo, Tokyo 113-0034, Japan
Telephone: +81-3-3818-4421 Telefax: +81-3-3818-4423



Head office·factory

2, December, 2013

Declaration of Conformity

We declare, under our solo responsibility, that the products:

Product Name	SHIN-NIPPON Chart Projector
Model No.	CP-500
Classification	Class I

The requirements of Council Directive:

93/42/EEC — concerning medical devices, Annex VII

Applied Harmonized Standards:


EN ISO13485:2003	Quality management systems – system requirement for regulation purpose
EN ISO15004-1:2009	Ophthalmic instruments – Fundamental requirements and test methods
EN60601-1:2006	Medical electrical equipment – Part 1: General requirement for safety
EN60601-1-2:2007	Medical electrical equipment – Part 1-2: General requirement for safety – Collateral standard: Electromagnetic compatibility – requirements and tests
ISO 8596:2009	Ophthalmic optics – Visual acuity testing – Standard optotype and its presentation
ISO 10938:1998	Ophthalmic instruments– Chart Projectors
JIS T7310 : 2002	Chart projectors

Manufacturer's name: TOWA MEDICAL INSTRUMENTS Co., Ltd.

1698-2, Oaza-Nakano, Nakanoshi, Nagano 383-0013, Japan

The authorized signatory to this declaration:

Signature:



Name: Koichi Kasuya

Position: President